



www.lcndt.net

We can do better



ISO9001 Registered

PRODUCT CERTIFICATE

PRODUCT Densitometer

MODEL DT-300

SERIAL NO. 300112

MEASUREMENT BASED ON ISO 14807:2001 Photography
 -- Transmission and reflection densitometers
 -- Method for determining performance

MEASUREMENT RESULT QC Passed

LED lighted table:

USB Charger:

Calibration Report:

SAMPLE CERTIFICATE

AGFA Density Strip (SN.: <u>4502200</u>)								
Num.	0	1	2	3	4	5	6	7
Density:	0.15	0.30	0.60	0.90	1.21	1.50	1.80	2.10
Cal.:	0.16	0.31	0.61	0.90	1.22	1.51	1.79	2.09
AGFA Density Strip								
Num.	8	9	10	11	12	13	14	
Density:	2.41	2.69	2.98	3.26	3.59	3.91	4.21	
Cal.:	2.42	2.68	2.97	3.27	3.59	3.92	4.22	

Effective Seal

INSPECTOR _____

MANAGER _____

Date of Measurement: _____

Lu Cheng NDT Equipment Corp.

E-mail: Sales@Lcndt.net (For sales) gm@Lcndt.net (For service)

[Http://www.Lcndt.net](http://www.Lcndt.net)